

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>TBS Puffer (10x), steril</b>	<b>Kat.Nr./Cat.No. : 42596</b>
TBS Buffer (10x), sterile	<b>Lot/Contr.No. : P240351</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare, farblose Lösung clear, colourless solution	entspricht corresponds
<b>pH</b>	1:10 verdünnt 1:10 diluted	7.2 – 7.6	7.2
<b>Steriltest</b> sterile test		kein Wachstum no growth	entspricht corresponds
<b>Zusammensetzung</b> composition			
NaCl (# 30183)		1.5 M	
KCl (# 26868)		30 mM	
Tris (# 37180)		250 mM	
<b>Mindesthaltbarkeit</b> minimum shelf life			13.11.2026
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

<p><b>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.</b>  <b>Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.</b>          We do not guarantee that the product can be used for a special application.          This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 28.11.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.