

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

|                                 |                         |               |
|---------------------------------|-------------------------|---------------|
| <b>SERVALight PreMix Helios</b> | <b>Kat.Nr./Cat.No.:</b> | <b>42657</b>  |
| CL HRP WB Substrate             | <b>Lot/Contr.No.:</b>   | <b>240455</b> |

| <b>Parameter</b><br>parameter                   | <b>Methode</b><br>method              | <b>Spezifikation</b><br>specification         | <b>Ergebnis</b><br>result |
|---|---------------------------------------|---|---------------------------|
| <b>Aussehen</b><br>appearance                   |                                       | klare, gelbe Lösung<br>clear, yellow solution | entspricht<br>corresponds |
| <b>Anwendung</b><br>application                 | Chemilumineszenz<br>chemiluminescence | entspricht<br>corresponds                     | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                                       |   | 22.05.2025                |
| <b>Lagerung (°C)</b><br>storage                 |                                       |   | +15 bis/ to +30           |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 27.05.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.