

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

NetFix[®] for PAG
265 x 125 mm

Kat.Nr./Cat.No. : 42775

Lot/Contr.No. : P190152

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Größe (mm) size		265 x 125	
Fadendicke/ Durchmesser Thread thickness/ Yarn diameter		40 µm	entspricht corresponds
Anwendungstest tested for use	Haftungstest adhesion test	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			05.04.2022
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.07.2019

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.