

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

SERVALYT® PRECOTES® pH 4-6	Kat.Nr./Cat.No. : 42875
300 µm, 125x125 mm	Lot/Contr.No. : P160026

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Beschreibung description	Gebrauchsfertige Polyacrylamidgele mit Servalyt, die an eine Polyesterfolie gebunden sind. Gelformat: 125x125 mm, 300 µm Ready-to-use polyacrylamide gels with servalyte bound to polyester sheet. Gel format: 125x125 mm, 300 µm		
pH-Servalyt	2% in H ₂ O, RT	5.20 ± 0.4	5.1
Leitfähigkeit (mS/cm) conductivity	2% in H ₂ O, RT	0.150 - 0.350	0.256
Proteintrennung protein separation	IEF, 300 µm-Gel, 5 °C, 3500 Vh	entspricht Standard conform reference	entspricht conform
Haftung des Gels adhesion of the gel	IEF, 300 µm-Gel, 5 °C, 3500 Vh	Haftung ja adhesion yes	ja yes
Mindesthaltbarkeit minimum shelf life			18.07.2018
Lagerung (°C) storage			+2 bis/ to +8

<p>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert. Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle. We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 28.04.2017

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.