

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Anodenflüssigkeit 3 für IEF</b>	<b>Kat.Nr./Cat.No. : 42984</b>
Anode fluid 3 for IEF	<b>Lot/Contr.No. : P140102</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		klare, farblose Lösung clear, colorless solution	entspricht corresponds
<b>Inhalt (in 50 ml)</b> content		0.17g L-Asparaginsäure 0.18g L-Glutaminsäure 0.17g L-aspartic acid 0.18g L-glutamic acid	
<b>pH</b>		2.8 – 3.8	3.0
<b>Osmolarität (mOsmol/l)</b> osmolarity		40 – 70	49
<b>Leitfähigkeit (µS/cm)</b> conductivity		< 1000	487
<b>Anwendung (IEF)</b> use		entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			02/ 2015
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

<p><b>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.</b>  <b>Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.</b>          We do not guarantee that the product can be used for a special application.          This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 18.02.2014**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.