

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Applikatorstreifen 7 x 1</b> Applicator strips 7 x 1	<b>Kat.Nr./Cat.No. : 42989</b> <b>Lot/Contr.No. : 200766</b>
--	---

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Länge</b> length	263 x 6 x 1 mm  min. 24 Probenschlitze/ sample slots: 7 x 1 mm	
<b>Anwendung</b> application	Probenauftrag sample application	
<b>Mindesthaltbarkeit</b> minimum shelf life		28.07.2025
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 29.07.2020**

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.