

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

Applikatorstreifen 7 x 1	Kat.Nr./Cat.No. : 42989
Applicator strips 7 x 1	Lot/Contr.No. : 210877

Parameter parameter	Spezifikation specification	Ergebnis result
Länge length	263 x 6 x 1 mm min. 24 Probenschlitze/ sample slots: 7 x 1 mm	
Anwendung application	Probenauftrag sample application	
Mindesthaltbarkeit minimum shelf life		21.07.2026
Lagerung (°C) storage		+15 bis/ to +30

SERVA Electrophoresis GmbH

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 17.09.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.