

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

Applikatorstreifen 7 x 1 Applicator strips 7 x 1	Kat.Nr./Cat.No. : 42989 Lot/Contr.No. : 250315
--	---

Parameter parameter	Spezifikation specification	Ergebnis result
Länge length	263 x 6 x 1 mm	entspricht corresponds
	min. 24 Probenschlitze/ sample slots: 7 x 1 mm	entspricht corresponds
Anwendung application	Probenauftrag sample application	
Mindesthaltbarkeit minimum shelf life		28.03.2030
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 02.04.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.