

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

| | |
|---|---------------------------------|
| SERVAGel(TM) TG PRiME(TM) 14 % | Kat.Nr./Cat.No. : 43270 |
| 10 cm x 10 cm, 1 mm, 10 sample wells | Lot/Contr.No. : V23087-1 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|---|---|--|
| Beschreibung description | Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 10 Probetaschen Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm, 10 sample wells | | |
| Testbedingungen testing conditions | Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht. Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel. | | |
| Proteintrennung protein separation | Albumin (BSA) Carbonic anhydrase Trypsin Inhibitor (soy bean) | 67 kDa entspricht/conform 29 kDa entspricht/conform 21 kDa entspricht/conform | entspricht/conform entspricht/conform entspricht/conform |
| Aussehen appearance | Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels. Protein separation shows sharp and straight bands over the whole separation range of the gel. | | |
| Mindesthaltbarkeit minimum shelf life | | | 18.08.2024 |
| Lagerung (°C) storage | | | +2 bis/ to +8 |

| |
|---|
| <p>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert. Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle. We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.</p> |
|---|

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 20.07.2023

Dr. Judith Koch

Marina Konter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.