

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS



<b>SERVAGel(TM) TG PRiME(TM) 4 - 12 %</b> <b>10 cm x 10 cm, 1 mm, 10 sample wells</b>	<b>Kat.Nr./Cat.No. : 43274</b> <b>Lot/Contr.No. : V24020-1</b>
--	---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Beschreibung</b> description	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 10 Probestaschen  Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x 10 cm, Thickness: 1 mm, 10 sample wells		
<b>Testbedingungen</b> testing conditions	Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht.  Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel.		
<b>Proteintrennung</b> protein separation	Albumin (BSA) 67 kDa Carbonic anhydrase 29 kDa Trypsin inhibitor (soy bean) 21 kDa	entspricht/conform entspricht/conform entspricht/conform	entspricht/conform entspricht/conform entspricht/conform
<b>Aussehen</b> appearance	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.  Protein separation shows sharp and straight bands over the whole separation range of the gel.		
<b>Mindesthaltbarkeit</b> minimum shelf life			14.03.2025
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 16.02.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.