

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

**SERVAGel(TM) TG PRiME(TM) 8 - 16 %**  
**10 cm x 10 cm, 1 mm, 15 sample wells**

**Kat.Nr./Cat.No.:** 43290  
**Lot/Contr.No. :** V24097-1

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result	
<b>Beschreibung</b> description	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 10 Probetaschen  Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm, 10 sample wells			
<b>Testbedingungen</b> testing conditions	Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht.  Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel.			
<b>Proteintrennung</b> protein separation	Albumin (BSA)	67 kDa	entspricht/conform	entspricht/conform
	Carbonic anhydrase	29 kDa	entspricht/conform	entspricht/conform
	Trypsin inhibitor (soy bean)	21 kDa	entspricht/conform	entspricht/conform
<b>Aussehen</b> appearance	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.  Protein separation shows sharp and straight bands over the whole separation range of the gel.			
<b>Mindesthaltbarkeit</b> minimum shelf life	17.02.2025			
<b>Lagerung (°C)</b> storage	+2 bis/ to +8			

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**

**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 22.07.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**

This report has been computer generated and does not contain a signature.