

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS



**SDS Gels 15% 52S**

**Size: 250 mm x 125 mm x 0.45 mm**

**Kat.Nr./Cat.No. : 43853**

**Lot/Contr.No. : 240239**

| Parameter<br>parameter                          | Methode<br>method                                                                                                                                                                 | Spezifikation<br>specification | Ergebnis<br>result     |
|-------------------------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------|------------------------|
| <b>Beschreibung</b>                             | Gebrauchsfertige Polyacrylamid-Gele, die an eine Kunststoffolie gebunden sind, zur Verwendung in der SDS-Gelelektrophorese von Proteinen.<br>Format: 250 x 125 mm, Dicke: 0.45 mm |                                |                        |
| description                                     | Ready-to-use polyacrylamide-gels, bound to a plastic foil, for use in SDS gel electrophoresis of proteins.<br>Format: 250 x 125 mm, Thickness: 0.45 mm                            |                                |                        |
| <b>Testbedingungen</b>                          | Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht.                                                            |                                |                        |
| testing conditions                              | Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel.                                                |                                |                        |
| <b>Proteintrennung</b><br>protein separation    | SDS-Gelelektrophorese<br>SDS gel electrophoresis                                                                                                                                  | entspricht<br>conforms         | entspricht<br>conforms |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life | 04.03.2026                                                                                                                                                                        |                                |                        |
| <b>Lagerung (°C)</b><br>storage                 | +2 bis/ to +8                                                                                                                                                                     |                                |                        |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**

**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 11.03.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**

This report has been computer generated and does not contain a signature.

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                                   |                         |               |
|-----------------------------------|-------------------------|---------------|
| <b>Buffer Kit for 1D SDS Gels</b> | <b>Kat.Nr./Cat.No.:</b> | <b>43383</b>  |
|                                   | <b>Lot/Contr.No.:</b>   | <b>250085</b> |

| <b>Parameter</b><br>parameter                   | <b>Methode</b><br>method                                                                                                                                                                                                                                                                                                 | <b>Spezifikation</b><br>specification | <b>Ergebnis</b><br>result |
|-------------------------------------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------|---------------------------|
| <b>Beschreibung</b><br>description              | Kit enthält SDS-Anodenpuffer, SDS-Kathodenpuffer, SDS-Probenpuffer neutral 2X und SDS-Probenverdünnungspuffer neutral 1X sowie Kühlkontaktflüssigkeit.<br><br>Kit contains SDS anode buffer, SDS cathode buffer and SDS buffer neutral sample buffer 2X, SDS buffer neutral sample diluter 1X and cooling contact fluid. |                                       |                           |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                                                                                                                                                                                                                                                                                                                          |                                       | 04.07.2026                |
| <b>Lagerung (°C)</b><br>storage                 |                                                                                                                                                                                                                                                                                                                          |                                       | +15 bis/ to +30           |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 28.01.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.