

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Cooling Contact Fluid</b>	<b>Kat.Nr./Cat.No.:</b>	<b>43371</b>
	<b>Lot/Contr.No.:</b>	<b>P250076</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Anwendung</b> application	für den Gebrauch in der horizontalen Elektrophorese for use in horizontal electrophoresis		
<b>Mindesthaltbarkeit</b> minimum shelf life			07.02.2028
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

SERVA Electrophoresis GmbH

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 06.03.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.