

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

SERVA Lightning Sci2	Kat.Nr./Cat.No. :	43404
	Lot/Contr.No. :	230654

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
monoisotope Masse (Da) monoisotopic mass		550.20	
Aussehen appearance		gelbes Pulver yellow powder	entspricht corresponds
Anregung (nm) excitation		490	
Emission (nm)		510	
Anwendungstest application test	für 2D PAGE for 2D PAGE	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			05.06.2025
Lagerung (°C) storage			-15 bis / to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 22.05.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.