

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>SERVA Lightning Sci2</b>	<b>Kat.Nr./Cat.No. :</b>	<b>43404</b>
	<b>Lot/Contr.No. :</b>	<b>240827</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>monoisotope Masse (Da)</b> monoisotopic mass		550.20	
<b>Aussehen</b> appearance		gelbes Pulver yellow powder	entspricht corresponds
<b>Anregung (nm)</b> excitation		490	
<b>Emission (nm)</b>		510	
<b>Anwendungstest</b> application test	für 2D PAGE for 2D PAGE	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			16.09.2025
<b>Lagerung (°C)</b> storage			-15 bis / to -25

<p><b>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.</b>  <b>Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.</b>          We do not guarantee that the product can be used for a special application.          This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**
**Druckdatum/ printing date: 23.09.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.