

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | | |
|-----------------------------|--------------------------|---------------|
| SERVA Lightning Sci5 | Kat.Nr./Cat.No. : | 43406 |
| | Lot/Contr.No. : | 211189 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|--|----------------------------|---------------------------------------|---------------------------|
| monoisotope Masse (Da) monoisotopic mass | | 580.32 | |
| Aussehen appearance | | blaues Pulver blue powder | entspricht corresponds |
| Anregung (nm) excitation | | 645 | |
| Emission (nm) | | 660 | |
| Anwendungstest application test | für 2D PAGE for 2D PAGE | entspricht corresponds | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | | 12.10.2022 |
| Lagerung (°C) storage | | | -15 bis / to -25 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.10.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.