

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS



Visking® Dialysierschlauch	Kat.Nr./Cat.No. :	44110
Visking® dialysis tubing	Lot/Contr.No. :	241102

Parameter parameter	Spezifikation specification	Ergebnis result
Material material	regenerierte Cellulose regenerated cellulose	entspricht corresponds
Flachbreite (mm) flat width	25	entspricht corresponds
Durchmesser (mm) diameter	16	entspricht corresponds
Volumen/Länge (ml/cm) volume/length	2	entspricht corresponds
Ausschlußgrenze (Dalton) exclusion limit	12 000 – 14 000	entspricht corresponds
Porengröße (Å) pore size	ca. 25	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		16.12.2029
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 16.12.2024

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.