

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Serdolit[®] Blau p.A.	Kat.Nr./Cat.No. : 45520
Serdolit [®] Blue analytical grade	Lot/Contr.No. : 101144

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		blaues Korn blue particles	entspricht corresponds
Korngröße (mesh) particle size		16 – 50 (0,3 – 1.2mm)	entspricht corresponds
Kapazität (eq/l) capacity		min. 0.8	1.1
Trocknungsverlust (%) loss on drying	3h, 105°C	60 – 70	63.5
Gegenion counter ion		OH ⁻	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			10/ 2016
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 07.08.2015

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.