

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Ameisensäure 99% für LC-MS Formic acid 99% for LC-MS	Kat.Nr./Cat.No. : 45640 Lot/Contr.No. : 180312
--	---

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		HCOOH	
Molekulargewicht molecular weight		46.03	
Aussehen appearance		klare, farblose Flüssigkeit clear, colourless liquid	entspricht corresponds
Gehalt (%) assay	azidimetrisch acidimetric	min. 99.0	99.54
Brechungsindex refractive index	20 °C	1.3709 - 1.3719	1.3712
Abdampfrückstand (ppm) residue on evaporation		max. 10	5
Transmission (%)	260 nm 270 nm	min. 20 min. 85	36.44 96.33
Aluminium (ppm)	Al	max. 0.05	< 0.05
Magnesium (ppm)	Mg	max. 0.1	< 0.1
Eisen/ iron (ppm)	Fe	max. 0.2	< 0.2
Kalzium/ calcium (ppm)	Ca	max. 0.2	< 0.2
Kalium/ potassium (ppm)	K	max. 0.1	< 0.1
Natrium/ sodium (ppm)	Na	max. 0.5	< 0.5
Mindesthaltbarkeit minimum shelf life			30.06.2019
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.03.2018

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.