

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

|  |                                |
|--|--------------------------------|
| <b>Collagen R Lösung 0.2 %, steril</b> | <b>Kat.Nr./Cat.No. : 47254</b> |
| Collagen R solution 0.2 %, sterile     | <b>Lot/Contr.No. : 240343</b>  |

| <b>Parameter</b><br>parameter                   | <b>Spezifikation</b><br>specification                       | <b>Ergebnis</b><br>result |
|---|---|---------------------------|
| <b>Molekulargewicht</b><br>molecular weight     | ca. 300 000   |                           |
| <b>Aussehen</b><br>appearance                   | viskose Lösung<br>viscous solution                          | entspricht<br>corresponds |
| <b>Konzentration (%)</b><br>concentration       | 2 mg/ ml in Essigsäure 0.1%<br>2 mg/ ml in acetic acid 0.1% | entspricht<br>corresponds |
| <b>pH</b>                                       | 3.2 – 3.8   | 3.4                       |
| <b>Steriltest</b><br>sterile test               | kein Wachstum<br>no growth                                  | entspricht<br>corresponds |
| <b>Gelbildung</b><br>gel formation              | positiv<br>positive   | entspricht<br>corresponds |
| <b>Ursprung</b><br>origin                       | Rattenschwanz<br>rat tail                                   | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |   | 10.04.2027                |
| <b>Lagerung (°C)</b><br>storage                 |   | +2 bis/ to +8             |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 16.04.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.