

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | | |
|---|--------------------------|---------------|
| Mycorase-Lösung (50X) für Mykoplasmen-Entfernung | Kat.Nr./Cat.No. : | 47987 |
| Mycorase solution (50x) for mycoplasma removal | Lot/Contr.No. : | 170171 |

| Parameter parameter | Spezifikation specification | Ergebnis result |
|---|---|---------------------------|
| Aussehen appearance | gelbliche Flüssigkeit yellowish liquid | entspricht corresponds |
| pH | | 9.5 |
| Osmolalität (mOsmol/ kg) osmolality | | 275 |
| Mindesthaltbarkeit minimum shelf life | | 26.01.2019 |
| Lagerung (°C) storage | | -15 bis/ to -25 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 23.02.2017

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.