

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

G 418 Lösung 50 mg/ ml, sterilfiltriert	Kat.Nr./Cat.No. : 47995
G 418 solution 50 mg/ ml, steril filtered	Lot/Contr.No. : P250141

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose bis gelbliche Lösung clear, colorless to pale yellow solution	entspricht corresponds
Konzentration (mg/ ml in H₂O) concentration		50 mg G-418	entspricht corresponds
pH		4.0 – 6.0	4.9
Osmolarität (mOsmol/ l) osmolarity			197
Steriltest sterile test		kein Wachstum no growth	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			17.04.2027
Lagerung (°C) storage			-15 bis/ to -25

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
 Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 05.05.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.