

## PRODUCT INFORMATION

**ReadyLyzer 10 MWCO 3.5 kDa**

**Cat. No.: 44630**

### PRODUCT DESCRIPTION

ReadyLyzer are ready-to-use devices for quick and efficient dialysis consisting of a low-binding plastic tube with an ultra-pure dialysis membrane pre-installed and screw cap.

---

**Application** Desalting, buffer exchange, removal of small molecular impurities, sample concentration

---

#### Dialysis Procedure

- Fill the ReadyLyzer with 10 ml dH<sub>2</sub>O, incubate it at least for 5 min and empty the tube.
- 

**IMPORTANT:** Make sure that there is no dH<sub>2</sub>O leakage.

Absorption of water by the dry membrane will decrease the water level.

---

- Pipet sample (3 – 10 ml) into the ReadyLyzer tube and close the screw cap. Pipet small sample volumes (e.g. 3 ml) close to the inner membrane.
  - Place the ReadyLyzer in the supplied floating rack. Position the rack according to sample volume: ≤ 5 ml middle, > 5 ml top of the device
  - Transfer it into a beaker (on a magnetic stirrer) containing a large volume of the desired buffer (100- to 1000-fold of the sample volume).
  - Adjust the speed of the stir bar.
  - Low-molecular salts and buffers (Tris-HCl, NaCl) equilibrate within 3 hours. Equilibration times for viscous samples will be longer.
- 

**IMPORTANT:** Optimal equilibration times for the dialysis must be determined by the user.

---

- Change dialysis buffer as necessary, at least 2-3 times
  - After dialysis remove sample carefully using a pipette and transfer the sample into a clean tube.
- 

#### Concentration Procedure

- Pipet sample into the ReadyLyzer or use already dialyzed sample, place it in bench top tube rack and let sample evaporate e.g. by using a fan to speed up the process.
  - Check sample frequently to prevent complete evaporation.
- 

**IMPORTANT:** When evaporation water, small molecules (buffer salts, reducing agents etc.) will also be concentrated in the sample.

---

**Storage** Store at + 15 °C to + 30 °C

---

Ver. 07/15